

## **DIN stellt Normen für medizinische Ausrüstung bis auf Weiteres kostenlos bereit**

Um die Herausforderungen der Corona-Krise zu bekämpfen, stellt DIN in Absprache mit der Europäischen Kommission bis auf Weiteres verschiedene Normen für medizinische Ausrüstung kostenlos zur Verfügung. Ziel ist es, dem wachsenden Mangel an Schutzmasken, -handschuhen und weiteren Produkten zu begegnen, mit dem viele europäische Länder zu kämpfen haben. Mit der Bereitstellung der Normen soll Unternehmen geholfen werden, die ihre Produktlinien umstellen wollen, um die so dringend benötigte Ausrüstung kurzfristig herzustellen. Der Einsatz von Normen kann dabei unterstützen, die Sicherheit von medizinischen Geräten und persönlicher Schutzausrüstung zu gewährleisten und Herstellern bei der Produktion praktische Hilfestellung zu geben. Alle kostenfreien Normen für medizinische Ausrüstung sind hier zu finden: [www.beuth.de/go/kostenlose-normen](http://www.beuth.de/go/kostenlose-normen)

## **Warum sind Normen üblicherweise kostenpflichtig?**

Normen sind grundsätzlich kostenpflichtige Produkte. Ihre Anwender profitieren von einer Reihe an Vorteilen: unter anderem Effizienzsteigerung, Kosteneinsparung, erleichterter Marktzugang, Senkung des Produkthaftungsrisikos. In Deutschland ist die Normung Selbstverwaltungsaufgabe der Wirtschaft. Dadurch ist gewährleistet, dass die Anwender von Normen selbst entscheiden, welche Normen marktgerecht sind. Die Anwender sorgen durch den Kauf von Normen dafür, dass die unabhängige, privatwirtschaftliche und effiziente Organisation der Normungsarbeit durch DIN als gemeinnützigem Verein erhalten bleibt.

## **Wer erstellt Normen?**

Normen entwickeln diejenigen, die sie später anwenden. Damit der Markt die Normen akzeptiert, sind eine breite Beteiligung, Transparenz und Konsens Grundprinzipien bei DIN. Jeder kann einen Antrag auf Normung stellen. Alle an einem Thema interessierten Kreise erhalten die Möglichkeit, mitzuwirken und ihre Expertise einzubringen. Vor der Verabschiedung werden die Norm-Entwürfe öffentlich gemacht und zur Diskussion gestellt. Die beteiligten Experten müssen sich über die endgültigen Inhalte grundsätzlich einig sein. Spätestens alle fünf Jahre werden Normen auf den Stand der Technik hin überprüft.

Weitere Informationen unter [www.din.de](http://www.din.de).

DIN EN 13795-1

DIN

ICS 11.140

Together with  
DIN EN 13795-2:2019-06,  
supersedes  
DIN EN 13795:2013-10

**Surgical clothing and drapes –  
Requirements and test methods –  
Part 1: Surgical drapes and gowns;  
English version EN 13795-1:2019,  
English translation of DIN EN 13795-1:2019-06**

Operationskleidung und -abdecktücher –  
Anforderungen und Prüfverfahren –  
Teil 1: Operationsabdecktücher und -mäntel;  
Englische Fassung EN 13795-1:2019,  
Englische Übersetzung von DIN EN 13795-1:2019-06

Vêtements et champs chirurgicaux –  
Exigences et méthodes d'essai –  
Partie 1: Champs et casaques chirurgicaux;  
Version anglaise EN 13795-1:2019,  
Traduction anglaise de DIN EN 13795-1:2019-06

Document comprises 37 pages

Um die Herausforderungen der Corona-Krise zu bekämpfen, stellt DIN diese Fassung der Norm bis auf Weiteres kostenlos bereit.  
Weitere aktuell kostenlose Normen für medizinische Ausrüstung unter [www.beuth.de/go/kostenlose-normen](http://www.beuth.de/go/kostenlose-normen).

Translation by DIN-Sprachendienst.

In case of doubt, the German-language original shall be considered authoritative.

*A comma is used as the decimal marker.*

## National foreword

This document (EN 13795-1:2019) has been prepared by Technical Committee CEN/TC 205 “Non-active medical devices” (Secretariat: DIN, Germany) with the active participation of German experts.

The responsible German body involved in its preparation was *DIN-Normenausschuss Medizin* (DIN Standards Committee Medicine), Working Committee NA 063-01-08 AA “Surgical textiles”.

This standard contains specifications giving detail to the essential requirements set out in EC Directive 93/42/EEC on medical devices.

In December 2008, the European Working Group CEN/TC 205/WG 14 decided to adapt the existing three parts of the standards series EN 13795 to the state of the art and thus to revise them, particularly with regard to the test methods. The revision was supposed to be divided into two phases. In the first phase, the three parts of the standards series EN 13795 were largely editorially combined. This phase was completed with the publication of EN 13795:2011. Due to necessary amendments of Annex ZA reflecting the relationship between this standard and the Medical Devices Directive (Directive 93/42/EEC), a revised new edition was published as EN 13795:2011+A1:2013 in 2013.

This document marks the end of the second phase, with the complete technical revision of the standard.

In addition, during the second phase of the revision of EN 13795, the decision was made to move the product “clean-air suit” to a separate part. The reason for this is that a clean-air suit is to meet different performance requirements than other surgical textiles.

## Amendments

This standard differs from DIN EN 13795:2013-10 as follows:

- a) the product “clean-air suit” has been moved to Part 2 of the EN 13795 standards series due to distinctive requirements and test methods;
- b) the title of the standard and the Scope have been adapted;
- c) normative references and the Bibliography have been updated;
- d) the clause “Terms and definitions” has been adapted;
- e) the performance requirements in Table 1 and Table 2 have been updated, in particular with regard to microbial penetration (wet);
- f) the former Clause 5 “Testing” has been moved to A.1 and has been editorially revised;
- g) the clause “Manufacturing and processing requirements and documentation” has been updated by adding documentary requirements and, in particular, by including a clause on the introduction of a QM system;
- h) the original Clause 7 “Information to be supplied by the manufacturer or processor” has been extended and its structure has been improved;

- i) the original Annex A, which included details on significant changes compared with the previous edition and consisted of three parts, has been deleted;
- j) Annex A “Testing” (former Annex B “Test methods”) has been completely revised and extended;
- k) a new Annex B “Rationales” has been added which provides precise reasons for the important requirements of this standard and is particularly intended for use by those who are familiar with the subject of this standard but who have not participated in its development;
- l) the original Annex C “Prevention of infection in the operating room” has been deleted;
- m) Annex C (former Annex D) “Information on further characteristics” has been revised and extended; e.g. subclauses on “Flammability” and “Electrostatic discharge” have been included;
- n) a new Annex D “Environmental aspects” has been added;
- o) a new Annex E “Guidance to users for selecting products” has been added;
- p) Annex ZA on the relationship between this standard and the Medical Devices Directive (93/42/EEC) has been updated;
- q) the standard has been completely editorially revised.

#### Previous editions

DIN EN 13795-1: 2003-02, 2009-10  
DIN EN 13795-2: 2005-02, 2009-12  
DIN EN 13795-3: 2006-08, 2009-12  
DIN EN 13795: 2011-06, 2013-10

— This page is intentionally blank —

Befristete Sonderausgabe

EUROPEAN STANDARD

EN 13795-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2019

ICS 11.140

Supersedes EN 13795:2011+A1:2013

English Version

## Surgical clothing and drapes - Requirements and test methods - Part 1: Surgical drapes and gowns

Vêtements et champs chirurgicaux - Exigences et méthodes d'essai - Partie 1: Champs et casaque chirurgicaux

Operationskleidung und -abdecktücher - Anforderungen und Prüfverfahren - Teil 1: Operationsabdecktücher und -mäntel

This European Standard was approved by CEN on 24 October 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

© 2019 CEN All rights of exploitation in any form and by any means reserved worldwide for CEN national Members.

Ref. No. EN 13795-1:2019 E

Um die Herausforderungen der Corona-Krise zu bekämpfen, stellt DIN diese Fassung der Norm bis auf Weiteres kostenlos bereit.  
Weitere aktuell kostenlose Normen für medizinische Ausrüstung unter [www.beuth.de/go/kostenlose-normen](http://www.beuth.de/go/kostenlose-normen).

## Contents

	Page
European foreword.....	4
Introduction .....	6
1 Scope.....	7
2 Normative references.....	7
3 Terms and definitions .....	8
4 Performance requirements.....	11
5 Manufacturing and processing requirements and documentation.....	13
6 Information to be supplied with the product .....	13
6.1 Information to be supplied to the user .....	13
6.2 Information to be supplied to the processor .....	14
Annex A (normative) Testing.....	15
A.1 General.....	15
A.2 Test methods and conformance.....	15
A.2.1 Test method for evaluation of cleanliness microbial/bioburden .....	15
A.2.2 Test method for evaluation of particle release .....	16
A.2.3 Test method for evaluation of liquid penetration.....	16
A.2.4 Test method for evaluation of bursting strength in dry and wet state.....	16
A.2.5 Test method for evaluation of tensile strength in dry and wet state .....	17
A.2.6 Test method for evaluation of dry microbial penetration .....	17
A.2.7 Test method for evaluation of wet microbial penetration.....	17
A.2.8 Test method for evaluation of biocompatibility .....	17
A.3 Treatment of results .....	18
Annex B (informative) Rationales.....	19
B.1 General.....	19
B.2 Cleanliness - microbial .....	19
B.3 Particle release .....	19
B.4 Resistance to liquid penetration .....	20
B.5 Bursting strength - dry and wet .....	21
B.6 Tensile strength - dry and wet.....	21
B.7 Resistance to microbial penetration - dry.....	21
B.8 Resistance to microbial penetration - wet .....	23
B.9 Labelling.....	23
B.10 Treatment of results .....	24

<b>Annex C (informative) Information on further characteristics .....</b>	<b>25</b>
<b>C.1 Comfort.....</b>	<b>25</b>
<b>C.2 Adhesion for fixation for the purpose of wound isolation.....</b>	<b>25</b>
<b>C.3 Liquid control.....</b>	<b>25</b>
<b>C.4 Flammability .....</b>	<b>26</b>
<b>C.5 Electrostatic discharge .....</b>	<b>26</b>
<b>Annex D (informative) Environmental aspects.....</b>	<b>27</b>
<b>Annex E (informative) Guidance to users for selecting products.....</b>	<b>28</b>
<b>E.1 Performance levels .....</b>	<b>28</b>
<b>E.2 Functional design.....</b>	<b>28</b>
<b>E.2.1 General .....</b>	<b>28</b>
<b>E.2.2 Critical and less critical areas .....</b>	<b>28</b>
<b>E.2.3 Size .....</b>	<b>29</b>
<b>E.2.4 Accessories.....</b>	<b>29</b>
<b>E.2.5 Comfort.....</b>	<b>29</b>
<b>E.2.5.1 General .....</b>	<b>29</b>
<b>E.2.5.2 Surgical gowns .....</b>	<b>29</b>
<b>E.2.5.3 Surgical drapes .....</b>	<b>30</b>
<b>E.3 Practical trials.....</b>	<b>30</b>
<b>Annex ZA (informative) Relationship between this European standard and the essential requirements of Directive 93/42/EEC [1993 OJ L 169] aimed to be covered .....</b>	<b>31</b>
<b>Bibliography .....</b>	<b>32</b>



## **European foreword**

This document (EN 13795-1:2019) has been prepared by Technical Committee CEN/TC 205 “Non-active medical devices”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2019, and conflicting national standards shall be withdrawn at the latest by October 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Together with EN 13795-2:2019, this document supersedes EN 13795:2011+A1:2013.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

EN 13795 consists of the following parts, under the general title *Surgical clothing and drapes — Requirements and test methods*:

- *Part 1: Surgical drapes and gowns*
- *Part 2: Clean air suits*

The following changes have been introduced:

- a) The product ‘clean-air suit’ has been moved to Part 2 of the EN 13795 standard series because of distinctive requirements and test methods;
- b) Alignment of the document title and the Scope;
- c) Revision of the Normative references and the Bibliography;
- d) Alignment of the Clause ‘Terms and definitions’;
- e) Review of the performance requirements in Table 1 and Table 2 especially with regard to ‘Cleanliness - Particulate matter’ and ‘Linting’, which have been combined as ‘Particle release’;
- f) Movement of former Clause 5 ‘Testing’ to A.1 and editorial alignment;
- g) Revision of Clause ‘Manufacturing and processing requirements’ by adding of documentary requirements and a section for the introduction of a QM system;
- h) Enhancement and improved structuring of Clause ‘Information to be supplied by the manufacturer or processor’;
- i) Deletion of the former Annex A ‘Details of significant changes between this document and the previous edition’ which consisted of 3 parts;
- j) Complete revision and extension of Annex A ‘Testing’ (formerly Annex B ‘Test methods’);

- k) Inclusion of a new Annex B 'Rationales' which provides precise reasons for the essential requirements of this document and which is intended for users aware of the subject of this document, but who have not participated in its development;
- l) Deletion of the former Annex C 'Prevention of infection in the operating room';
- m) Revision and extension of Annex C (formerly Annex D) 'Information on further characteristics'; e.g. inclusion of a Clause on 'Flammability' and 'Electrostatic discharge';
- n) Inclusion of a new Annex D 'Environmental aspects';
- o) Inclusion of a new Annex E 'Guidance to users for selecting products';
- p) Revision of Annex ZA on the relationship to the Medical Device Directive (93/42/EEC);
- q) Complete editorial revision.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## **Introduction**

The transmission of infective agents during invasive surgical procedures can occur in several ways (see informative Annex B).

Surgical drapes, including the intended use as a sterile field, and surgical gowns are used to minimize the spread of infective agents to and from patients' operating wounds, thereby helping to prevent post-operative wound infections (see Annex B).

The performance required of coverings for patients, clinical staff and equipment varies with, for example, the type and duration of the procedure, the degree of wetness of the operation field, the degree of mechanical stress on the materials and the susceptibility of the patient to infection.

The use of surgical gowns with resistance to the penetration of liquids can also diminish the risk to the operating staff from infective agents carried in blood or body fluids.

This document is intended to assist the communication between manufacturers and third parties with regard to material or product characteristics and performance requirements.

Therefore, Annex B provides comprehensive information on characteristics, measurement of performance and performance requirements. Annex C clarifies that this document does not include environmental provisions. Annex D provides information on characteristics regarded relevant in context with surgical gowns and drapes, however but not covered normatively (i.e. without applicable performance requirements). Annex E explains the concept of performance levels and provides guidance to users for selecting products.

This document focuses on Essential Requirements arising from the Medical Device Directive 93/42/EEC, which are applicable to surgical drapes and gowns. The requirements and guidance in this document are expected to be of help to manufacturers and users when designing, processing, assessing and selecting products. It is the intention of this document to ensure the same level of safety from single-use and reusable surgical clothing and drapes throughout their useful life.

Surgical gowns are used to minimize the transmission of infective agents between patients and clinical staff during surgical and other invasive procedures. Hereby, surgical gowns contribute to the clinical condition and the safety of patients as well as to the safety and health of users following up essential requirement 1 of Directive 93/42/EEC on Medical Devices. This document addresses the same level of protection for patients and users (i.e. the surgical team) by not differentiating the performance requirements for surgical gowns respectively. However, this document does not formally address any basic health and safety requirements of the Directive 89/686/EEC or Regulation (EU) 2016/425 on Personal Protective Equipment and does not provide specific guidance for surgical gowns intended by the manufacturer for dual use as medical device and personal protective equipment.

## 1 Scope

This document specifies information to be supplied to users and third party verifiers in addition to the usual labelling of medical devices (see EN 1041 and EN ISO 15223-1), concerning manufacturing and processing requirements. This document gives information on the characteristics of single-use and reusable surgical gowns and surgical drapes used as medical devices for patients, clinical staff and equipment, intended to prevent the transmission of infective agents between clinical staff and patients during surgical and other invasive procedures. This document specifies test methods for evaluating the identified characteristics of surgical drapes and gowns and sets performance requirements for these products.

This document does not cover requirements for resistance to penetration by laser radiation of products. Suitable test methods for resistance to penetration by laser radiation, together with an appropriate classification system, are given in EN ISO 11810.

This document does not cover requirements for incision drapes or films.

This document does not cover requirements for antimicrobial treatments for surgical gowns and drapes. Antimicrobial treatment can cause environmental risks such as resistance and pollution. However, antimicrobial treated surgical gowns and drapes fall under the scope of this document with respect to their use as surgical gowns and drapes.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 811:2018, *Textiles - Determination of resistance to water penetration - Hydrostatic pressure test (ISO 811:2018)*

EN 29073-3:1992, *Textiles - Test methods for nonwovens - Part 3: Determination of tensile strength and elongation*

EN ISO 139:2005,<sup>1</sup> *Textiles — Standard atmospheres for conditioning and testing (ISO 139:2005 + Amd. 1:2011)*

EN ISO 9073-10:2004, *Textiles - Test methods for nonwovens - Part 10: Lint and other particles generation in the dry state (ISO 9073-10:2003)*

EN ISO 10993-1:2009, *Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process (ISO 10993-1:2009)*

EN ISO 11737-1:2018, *Sterilization of medical devices — Microbiological methods — Part 1: Determination of a population of microorganisms on products (ISO 11737-1:2018)*

EN ISO 13938-1:1999, *Textiles - Bursting properties of fabrics - Part 1: Hydraulic method for determination of bursting strength and bursting distension (ISO 13938-1:1999)*

---

<sup>1</sup> Impacted by EN ISO 139:2005+A1:2011

EN ISO 22610:2006, *Surgical drapes, gowns and clean air suits, used as medical devices, for patients, clinical staff and equipment - Test method to determine the resistance to wet bacterial penetration (ISO 22610:2006)*

EN ISO 22612:2005, *Clothing for protection against infectious agents - Test method for resistance to dry microbial penetration (ISO 22612:2005)*

### **3 Terms and definitions**

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

#### **3.1 colony forming unit**

##### **CFU**

unit by which the culturable number of microorganisms is expressed

Note 1 to entry: The culturable number is the number of microorganisms, single cells or aggregates, able to form colonies on a solid nutrient medium.

#### **3.2 cleanliness**

freedom from unwanted foreign matter

Note 1 to entry: Such matter can be microorganisms, organic residues or particulate matter.

##### **3.2.1 cleanliness — microbial**

freedom from population of viable micro-organisms on a product and/or a package

Note 1 to entry: In practical use, microbial cleanliness is often referred to as 'bioburden'.

#### **3.3 critical product area**

product area with a greater probability to be involved in the transfer of infective agents to or from the wound, e.g. front and sleeves of surgical gowns

#### **3.4 infective agent**

micro-organism that has been shown to cause wound infections or that might cause infection in a member of the surgical team or the patient

#### **3.5 less critical product area**

product area less likely to be involved in the transfer of infective agents to or from the wound

### 3.6

#### **manufacturer**

natural or legal person with responsibility for the design, manufacture, packaging and labelling of a device before it is placed on the market under his own name, regardless of whether these operations are carried out by that person himself or on his behalf by a third party

Note 1 to entry: For more details, refer to the Medical Device Directive 93/42/EEC.

### 3.7

#### **particle release**

release of fibre fragments and other particles during mechanical stress simulating handling and use

### 3.8

#### **performance level**

discrete standard defined to classify products according to the performance requirements of this document

Note 1 to entry: With the introduction of two performance levels, this document acknowledges the fact that products are challenged to differing extents during surgical procedures, dependent upon the duration, mechanical stress and liquid challenge throughout the surgical procedure.

#### 3.8.1

##### **standard performance**

classification addressing minimum performance requirements for various characteristics of products used as medical devices in invasive surgical procedures

#### 3.8.2

##### **high performance**

classification addressing elevated performance requirements for various characteristics of products used as medical devices in invasive surgical procedures

Note 1 to entry: Examples of surgical procedures where elevated performance level should be considered are those where extensive exposure to liquid, mechanical stresses or longer surgical procedures can be expected.

### 3.9

#### **processor**

natural or legal person who processes products so that their performance complies with the requirements of this document

Note 1 to entry: A processor who places a product on the market is a manufacturer in the sense of this document.

Note 2 to entry: A processor of reusable products is often referred to as a 'reprocessor' and processing reusable products is often referred to as 'reprocessing' (as e.g. in Medical Device Directive 93/42/EEC). References in EN 13795-2 and this document to 'processors' include 'reprocessors' and to 'processing' include 'reprocessing'.

### 3.10

#### **product**

surgical gown, surgical drape including equipment covering

Note 1 to entry: In cases of surgical packs, each gown or drape is regarded as a product.

### 3.11

#### **resistance to liquid penetration**

ability of material to withstand the penetration of liquid(s) from one side of the material through to the other

### 3.12

#### **resistance to microbial penetration**

ability of material(s) to withstand penetration of micro-organisms from one side of the material through to the other

#### 3.12.1

##### **dry penetration**

effect of a combination of air movement and mechanical action by vibration on microbial penetration in dry condition

#### 3.12.2

##### **wet penetration**

effect of combination of wetness, pressure and rubbing on microbial penetration

### 3.13

#### **reusable product**

product intended by the manufacturer to be reprocessed and reused

### 3.14

#### **single-use product**

product intended to be used once only for a single patient

### 3.15

#### **sterile field**

area created by sterile surgical drape material where aseptic technique is practised

Note 1 to entry: A sterile field can be practised e.g. on a back table.

### 3.16

#### **surgical drape**

drape covering the patient or equipment to prevent transfer of infective agents

### 3.17

#### **surgical gown**

gown worn by a member of a surgical team to prevent transfer of infective agents

### 3.18

#### **surgical procedure**

surgical intervention performed by a surgical team

#### 3.18.1

##### **invasive surgical procedure**

surgical procedure penetrating skin or mucosa

## 4 Performance requirements

To comply with this document, products shall meet all the requirements specified in this document including Tables 1 or 2 (as appropriate to the product), when tested according to Annex A of this document throughout their useful life.

The biocompatibility of the product shall be evaluated and approved for acceptable risk.

If the manufacturer does not differentiate product areas, all areas shall meet the requirements for critical product areas.

If the intended purpose of a medical device specifies the use as a sterile field the requirements for surgical drapes and equipment covers apply as per Table 2.

For general information on testing and details on the test methods given in this clause including Tables 1 and 2 and their application for the purpose of this document, see Annex A.

NOTE 1 Performance requirements are specified depending on product area and performance level. However, for some characteristics the performance requirement will apply for all performance levels and product areas of the medical device.

NOTE 2 Information on characteristics, which cannot be properly evaluated (as 'adhesion for fixation for the purpose of wound isolation' or 'liquid control') or which are not regarded normative (as 'comfort') is given in Annex C.

**Table 1 — Characteristics to be evaluated and performance requirements for surgical gowns**

Characteristic	Test method (for normative references see Clause 2)	Unit	Requirement			
			Standard performance		High performance	
			Critical product area	Less critical product area	Critical product area	Less critical product area
Microbial penetration — Dry	EN ISO 22612	CFU	Not required	≤ 300 <sup>a</sup>	Not required	≤ 300 <sup>a</sup>
Microbial penetration — Wet	EN ISO 22610	$I_B$	≥ 2,8 <sup>b</sup>	Not required	6,0 <sup>b c</sup>	Not required
Cleanliness microbial / Bioburden	EN ISO 11737-1	CFU/100 cm <sup>2</sup>	≤ 300	≤ 300	≤ 300	≤ 300
Particle release	EN ISO 9073-10	log <sub>10</sub> (lint count)	≤ 4,0	≤ 4,0	≤ 4,0	≤ 4,0
Liquid penetration	EN ISO 811	cm H <sub>2</sub> O	≥ 20	≥ 10	≥ 100	≥ 10
Bursting strength — Dry	EN ISO 13938-1	kPa	≥ 40	≥ 40	≥ 40	≥ 40



Characteristic	Test method (for normative references see Clause 2)	Unit	Requirement			
			Standard performance		High performance	
			Critical product area	Less critical product area	Critical product area	Less critical product area
Bursting strength — Wet	EN ISO 13938-1	kPa	≥ 40	Not required	≥ 40	Not required
Tensile strength — Dry	EN 29073-3	N	≥ 20	≥ 20	≥ 20	≥ 20
Tensile strength — Wet	EN 29073-3	N	≥ 20	Not required	≥ 20	Not required

a Test conditions: challenge concentration 10<sup>8</sup> CFU/g talcum and 30 min vibration time.

b The Least Significant Difference (LSD) for  $I_B$  when estimated using EN ISO 22610, was found to be 0,98 at the 95 % confidence level. This is the minimum difference needed to distinguish between two materials thought to be different. Thus materials varying by up to 0,98  $I_B$  are probably not different; materials varying by more than 0,98  $I_B$  probably are different. (The 95 % confidence levels means that an observer would be correct 19 times out of 20 to accept these alternatives.)

c  $I_B = 6,0$  for the purpose of this document means: no penetration.  $I_B = 6,0$  is the maximum achievable value.

Table 2 — Characteristics to be evaluated and performance requirements for surgical drapes

Characteristic	Test method (for normative references see Clause 2)	Unit	Requirement			
			Standard performance		High performance	
			Critical product area	Less critical product area	Critical product area	Less critical product area
Microbial penetration — Dry	EN ISO 22612	CFU	Not required	≤ 300 <sup>a</sup>	Not required	≤ 300 <sup>a</sup>
Microbial penetration — Wet	EN ISO 22610	$I_B$	≥ 2,8 <sup>b</sup>	Not required	6,0 <sup>b c</sup>	Not required
Cleanliness microbial / Bioburden	EN ISO 11737-1	CFU/100 cm <sup>2</sup>	≤ 300	≤ 300	≤ 300	≤ 300
Particle release	EN ISO 9073-10	log <sub>10</sub> (lint count)	≤ 4,0	≤ 4,0	≤ 4,0	≤ 4,0
Liquid penetration	EN ISO 811	cm H <sub>2</sub> O	≥ 30	≥ 10	≥ 100	≥ 10
Bursting strength — Dry	EN ISO 13938-1	kPa	≥ 40	≥ 40	≥ 40	≥ 40

Characteristic	Test method (for normative references see Clause 2)	Unit	Requirement			
			Standard performance		High performance	
			Critical product area	Less critical product area	Critical product area	Less critical product area
Bursting strength — Wet	EN ISO 13938-1	kPa	≥ 40	Not required	≥ 40	Not required
Tensile strength — Dry	EN 29073-3	N	≥ 15	≥ 15	≥ 20	≥ 20
Tensile strength — Wet	EN 29073-3	N	≥ 15	Not required	≥ 20	Not required

- a Test conditions: challenge concentration  $10^8$  CFU/g talcum and 30 min vibration time.
- b The Least Significant Difference (LSD) for  $I_B$  when estimated using EN ISO 22610, was found to be 0,98 at the 95 % confidence level. This is the minimum difference needed to distinguish between two materials thought to be different. Thus materials varying by up to 0,98  $I_B$  are probably not different; materials varying by more than 0,98  $I_B$  probably are different. (The 95 % confidence level means that an observer would be correct 19 times out of 20 to accept these alternatives.)
- c  $I_B = 6,0$  for the purpose of this document means: no penetration.  $I_B = 6,0$  is the maximum achievable value.

## 5 Manufacturing and processing requirements and documentation

**5.1** The manufacturer and processor shall document that the requirements of this document are met and that the fitness for the intended purpose has been established for each use, both for single-use and reusable medical devices.

**5.2** The manufacturer/processor shall establish, document, implement and maintain a formal quality management system, which includes risk management and maintain its effectiveness. This quality management system shall include requirements throughout product realization, including development, design, manufacture, testing, packaging, labelling, distribution and, for reusable products, processing and life-cycle control.

Inputs for product realization shall include the outputs from risk management.

A quality system such as EN ISO 13485 is recommended, in case of processing of reusable products applied in accordance with EN 14065.

Packaging for terminally sterilized medical devices is recommended according to EN ISO 11607 series of standards.

For testing processes, quantitative physical, chemical and/or biological tests are preferred.

**5.3** A clinical evaluation for surgical drapes and gowns shall be carried out and shall consider the performance of the full draping and gowning system to establish fitness for purpose. The evaluation shall include the critical review of the applicable clinical literature and the results of post market surveillance and vigilance.

## 6 Information to be supplied with the product

### 6.1 Information to be supplied to the user

**6.1.1** In addition to the information to be supplied according to the Medical Device Directive 93/42/EEC, if the manufacturer or processor differentiates between critical and less critical areas of the product, he/she shall supply information to identify them.

**6.1.2** The following additional information shall be supplied on request:

- a) the identity or information on the test methods used;
- b) the results of testing and test conditions for the characteristics given in Clause 4.

**6.1.3** The manufacturer shall inform the user of residual risks due to any shortcomings of the protection measures adopted.

**6.1.4** The manufacturer shall provide sufficient information about intended use of the product or product system when conducting a surgical procedure. This shall include information on the performance level of the product.

**6.1.5** The manufacturer shall provide information on the flammability of the product and fire risks in relation with it on request.

## **6.2 Information to be supplied to the processor**

**6.2.1** For reusable products the manufacturer shall obtain information to be supplied to the processor on the number of reuses based on standardized processes, together with information on measures for maintaining the technical and functional safety of the medical device and packaging.

**6.2.2** For products to be terminally sterilized, the manufacturer shall supply instructions for the sterilization processes to be applied.

## Annex A (normative)

### Testing

#### A.1 General

**A.1.1** Testing for evaluation of the performance of products shall be done according to the test methods specified in A.2. All test results and test conditions shall be recorded and retained.

**A.1.2** Testing shall be performed on the finished product. If the product is to be used after sterilization, testing shall be performed on products after sterilization with the exception of microbial cleanliness. Testing shall include potential weak spots.

NOTE 1 Performance requirements can vary in relation to the areas of the product and the risk of involvement in the transfer of infective agents to or from the wound.

NOTE 2 To ensure product performance, combinations of materials or products in systems can be used.

NOTE 3 In particular, all types of joints in critical areas such as, e.g. seams in sleeves of surgical gowns, are regarded as potential weak spots.

**A.1.3** During manufacture and processing, testing shall be conducted according to the requirements of the manufacturer's and processor's quality system.

**A.1.4** Alternative test methods for monitoring may be used provided that they are validated and address the same characteristic and that the results have been shown to correlate with the test methods given in this document.

**A.1.5** Where the test methods of this document do not specify the atmosphere for pre-conditioning, conditioning and testing, the specifications of EN ISO 139 shall be applied. Prior to testing, the samples shall be conditioned in the relaxed state.

#### A.2 Test methods and conformance

##### A.2.1 Test method for evaluation of cleanliness microbial/bioburden

For evaluation of cleanliness — microbial, the product shall be tested according to EN ISO 11737-1.

NOTE EN ISO 11737-1 does not provide a fixed test method but specifies requirements for test methods and test mechanisms. The requirements of EN ISO 11737-1 are such that different test methods developed in accordance with it provide comparable results.

Five specimens shall be tested. The results shall be expressed as CFU/100 cm<sup>2</sup>. Report the individual results and determine  $M_d$  and  $U_q$  (see A.3).  $U_q$  shall be equal to or less than the performance requirements in Tables 1 and 2.

## **A.2.2 Test method for evaluation of particle release**

For evaluation of particle release, the product shall be tested according to EN ISO 9073-10.

NOTE 1 EN ISO 9073-10 allows for the test method to be conducted in a laminar flow hood. It is important to validate that laminar flow is occurring if equipment required for the test is located in the hood.

As specified in EN ISO 9073-10, ten specimens, five for each side of the material, shall be tested. The result of the test, i.e. the coefficient of linting, shall be calculated for particles in the size range 3 µm to 25 µm and reported as log<sub>10</sub> of the count value. Report the individual results and determine M<sub>d</sub> and U<sub>q</sub> (see A.3). U<sub>q</sub> shall be equal to or less than the performance requirements in Tables 1 and 2.

NOTE 2 Particles of this size range are considered to be capable of carrying microorganisms.

## **A.2.3 Test method for evaluation of liquid penetration**

For evaluation of liquid penetration, the product shall be tested according to EN ISO 811.

The following specific amendments to the procedure in EN ISO 811 apply for the purpose of this document:

- a) the test area shall be 100 cm<sup>2</sup>;
- b) the rate of increase of water pressure shall be (10 ± 0,5) cm/min;
- c) the temperature of the water shall be (20 ± 2) °C;
- d) the side of the product in contact with the test liquid shall be the outer side.

Five specimens shall be tested. Report the individual results and determine M<sub>d</sub> and L<sub>q</sub> (see A.3). L<sub>q</sub> shall be equal to or greater than the performance requirements in Tables 1 and 2.

As the test may be stopped once, the test limit hydrostatic pressure is exceeded or the measurement capability of the instrument is exceeded, the value to be used in the median and lower quartile calculations for hydrostatic pressure testing shall be the lower of the breakthrough number or the upper measurement capability if this has been exceeded.

## **A.2.4 Test method for evaluation of bursting strength in dry and wet state**

For evaluation of bursting strength, the product shall be tested according to EN ISO 13938-1. The size of the test area shall be 10 cm<sup>2</sup> (35,7 mm diameter). The preparation of samples for wet state testing shall be performed according to EN 29073-3.

The test conditions should be specified in the test report.

If there are differences in the test results of both sides of material, both sides should be tested and the results should be recorded.

Five specimens shall be tested. The pressure needed to break or compromise the barrier of the sample shall be reported. Report the individual results and determine M<sub>d</sub> and L<sub>q</sub> (see A.3). L<sub>q</sub> shall be equal to or greater than the performance requirements in Tables 1 and 2.

### A.2.5 Test method for evaluation of tensile strength in dry and wet state

For evaluation of tensile strength, the product shall be tested according to EN 29073-3 in the wet and dry states both in longitudinal and in lateral directions.

Five specimens shall be tested for each direction. The pressure needed to break or compromise the barrier of the sample shall be reported. Report the individual results and determine  $M_d$  and  $L_q$  (see A.3).  $L_q$  shall be equal to or greater than the performance requirements in Tables 1 and 2.

### A.2.6 Test method for evaluation of dry microbial penetration

For evaluation of dry microbial penetration, the product shall be tested according to EN ISO 22612.

If both sides of the material to be tested are different, the side intended to cover the contamination source during medical use as stated by the manufacturer shall be exposed to the inoculated donor in the test.

NOTE The side intended to cover the contamination source during medical use is, e.g. the inner side of a surgical gown or the patient or equipment side of a surgical drape.

If the product has an antimicrobial treatment, it shall be mentioned in the test report since it can influence the results.

Ten specimens shall be tested. Report the individual results and determine  $M_d$  and  $U_q$  (see A.3).  $U_q$  shall be equal to or less than the performance requirements in Tables 1 and 2.

### A.2.7 Test method for evaluation of wet microbial penetration

For evaluation of wet microbial penetration, the product shall be tested according to EN ISO 22610.

If both sides of the material to be tested are different, the side intended to cover the contamination source during medical use as stated by the manufacturer shall be exposed to the inoculated donor in the test.

NOTE The side intended to cover the contamination source during medical use is, e.g. the inner side of a surgical gown or the patient or equipment side of a surgical drape.

If the product has a known antimicrobial treatment, it shall be mentioned in the test report since it can influence the results.

Five specimens shall be tested. Report the results as per EN ISO 22610 including barrier index  $I_B$  as per EN ISO 22610:2006, C.4. The barrier index  $I_B$  shall be equal to or higher than the performance requirements in Tables 1 and 2.

### A.2.8 Test method for evaluation of biocompatibility

The manufacturer shall complete the evaluation of the surgical drape/gown according to EN ISO 10993-1:2009 and report the results of the evaluation.

### A.3 Treatment of results

In order to determine whether a sample conforms to the performance requirements of this document, it is necessary to convert the replicate results from a test into an acceptance value (or test statistic). The median ( $M_d$ ) was the chosen value (see Rationale Annex), together with one of two test statistics a) the lower quartile value ( $L_q$ ) for minimum performance ( $PR_{\min}$ ) and b) the upper quartile ( $U_q$ ) for maximum performance ( $PR_{\max}$ ).

For conformance of the product can be determined:

- $L_q \geq PR_{\min}$  (see Tables 1 and 2)
- $U_q \leq PR_{\max}$  (see Tables 1 and 2)
- $M_d$ ,  $L_q$  and  $U_q$  (or any percentile value)

by using the following general method.

To calculate the  $k^{\text{th}}$  percentile (where  $k$  is 25 for identifying the lower quartile number and 75 for identifying the upper quartile value):

1. Order all the values in the data set from the smallest to largest.
2. Multiply  $k$  percent by the total number of values,  $n$ . This product is called the index.
3. If the index obtained in step 2 is not a whole number, round it up to the nearest whole number and go to step 4a. If the index obtained in step 2 is a whole number, go to step 4b.
- 4a. Count the values in your data set from left to right (from the smallest to the largest value) until you reach the number indicated by step 3. The corresponding value in the data set is the  $k^{\text{th}}$  percentile.
- 4b. Count the values in your data set from left to right until you reach the number indicated by step 2. The  $k^{\text{th}}$  percentile is the average of that corresponding value in the data set and the value that directly follows it.

## Annex B (informative)

### Rationales

#### B.1 General

This annex provides a concise rationale for the important requirements of this document and is intended for use by those who are familiar with the subject of this document but who have not participated in its development. An understanding of the reasons for the main requirements is considered essential for its proper application. Furthermore, as clinical practices and technologies change, it is believed that rationales for the present requirements will facilitate any revisions of this document necessitated by those developments.

The first task undertaken by CEN/TC 205/WG 14 in its early days was deciding on the key product characteristics which needed to be assessed. After much consideration four categories emerged, namely barrier properties, strength properties relevant to maintaining barrier properties, particle release and bioburden level to ensure successful sterilization. Most of the performance limits in this document are based on expert consensus.

#### B.2 Cleanliness – microbial

The test for microbial cleanliness is intended to estimate the numbers of viable organisms on the products, **before** they are sterilized. This is frequently referred to as the 'bioburden', which manufacturers routinely measure, and use to determine the appropriate sterilization criteria for their products.

Note that this test is **not** a sterility test. In a bioburden (cleanliness) test, the presence of microorganisms is expected, and the test is designed to quantify the amount of microorganisms present (for example, through rinsing, filtering and counting). In a sterility test, the **absence** of microorganisms is expected, and a different methodology is used.

The cleanliness limit of 300 CFU (Tables 1 and 2) is based on the experience of manufacturers and what is routinely achievable at present. It is also a figure which industry state is acceptable to Notified Bodies as representing a bioburden capable of being dealt with by the sterilization methods available. Finally, it was also chosen as being a reasonable level for products which will not undergo a cleaning/disinfecting process prior to sterilization, such as single-use products.

The Working Group acknowledges that the device will usually have undergone a 'terminal sterilization' [18] process before clinical users receive it. Consequently, the requirements for cleanliness – microbial are set in anticipation of the sterilization process to be applied terminally.

#### B.3 Particle release

This method is designed to measure the release of particles from the device.

Particle release is a concern during surgery as foreign body contamination can cause an increased frequency of postoperative complications such as keloids, wound dehiscence, incisional hernias, chronic abscesses, intestinal obstruction and, in some circumstances, even death [19], [20]. Fibres from gowns and drapes which have been deposited in wounds have been shown to cause post-operative granulomas [21], [22]. Blood clots around fibres can cause emboli, obstructing vital blood vessels [23]. Fibres can also reduce the ability of tissue to resist infection, due to impaired function of the blood and tissue macrophage systems [24], [25].



As well as having a direct effect clinically, an indirect effect is observed, whereby fibres and particles released from operating room materials can deposit on surfaces in the operating room, providing a potential vector for microorganisms to be carried into wounds and cavities [26]. See section on "Resistance to microbial penetration" for a discussion on contamination versus infection.

In 1997, CEN/TC 205/WG 14 passed resolutions requiring both linting and cleanliness to be covered as normative parts of this document. Linting was defined as material created by mechanical handling of the material (such as flexing and rubbing during normal use), originating from the material itself. The 'ad hoc' Linting group in 1999 discussed a proposal that 'foreign matter' is expected to be released at the beginning of a flexion test, but linting (release of particles from the material itself) will occur throughout the testing. A method was proposed which gave an estimate of foreign particles and lint, and this proposal was accepted by CEN/TC 205/WG 14 in 1999. Documents from that period state that the first three time steps have significant peaks which are due to the foreign and loose matter, and subsequent counts are due to linting.

Thus the original version of this document included requirements for linting and particulate cleanliness, intended to differentiate loose particles from lint, and since there was no simple method, a cut-off point of 90 s was chosen based on examination of the graphs. Recently, CEN/TC 205/WG 14 has removed the requirement for particulate cleanliness from this document as it believes that the distinction between particulate cleanliness and linting was purely theoretical, with no evidence being presented to demonstrate that the original supposition of loose matter being released in the first 90 s was correct. Although there is no evidence that the theoretical concerns were unsubstantiated, it has been agreed that the performance characteristic which is of practical importance is total particles released from the material. Thus the new requirement is for a **total** particle release figure, which will also include loose particulate matter.

We do not believe that this will have any effect on the clinical acceptability or performance of the devices, as the amended test for 'Particle release' measures **all** the particles released during the test period which are thought to be clinically relevant.

The particulate size range of 3 µm to 25 µm has been chosen based on the opinion that particles smaller than 3 µm are too small to carry microorganisms, and particles larger than 25 µm are too large to remain airborne because of gravity. This is supported in work published by Noble in 1963 who found that "*Organisms associated with human disease or carriage were usually found on particles in the range 4 µm to 20 µm equivalent diameter*".

#### B.4 Resistance to liquid penetration

Also known as the 'hydrostatic head test', this test is a standard test used for textiles, which measures how high a column of water has to be before it penetrates through the material under test. It is generally accepted to be a measure of the water-resistant properties of a material.

It is relevant to surgical fabrics as it is related to the ability of the fabric to prevent splashes of fluid and droplets penetrating the fabric under mechanical pressure.

The limits of 10 cm, 20 cm, 30 cm and 100 cm H<sub>2</sub>O (Tables 1 and 2) are based on manufacturer experience with similar ranges of devices in the market place.

This particular test is based on water and that whilst CEN/TC 205/WG 14 is aware that these devices are exposed to other substances such as fats in the operating room, the water test is an established and well accepted test to characterize barrier fabrics by the textile industry.

The liquid penetration test is also acknowledged as a useful and simple test to monitor both single-use and reusable fabrics during processing and between uses, as performing wet bacterial barrier penetration tests routinely on batches is impractical.

EN ISO 811 allows for two different temperatures and two different rates of rise for testing. Both test conditions influence the test result and hence the evaluation of conformity with the requirements of this document. As a consequence, the temperature and rate of rise was specified in this document.

Based on the condition usually used for testing by laboratories and manufacturers the temperature has been specified as  $(20 \pm 2)$  °C.

As for the rate of rise members of WG 14 have undertaken tests of multiple materials. The analyses of this data show that a wider spread of results is seen in the results with the faster rate rise (60 cm/min), which implies less precision in the test results. In addition, when tested at 60 cm/min rise, the results are elevated compared to the results at 10 cm/min, and some materials considered unsatisfactory, which fail at 10 cm/min would pass at 60 cm/min. Therefore, to ensure consistency, a decision has been made by WG 14 to only allow a 10 cm/min rate of rise when testing to compliance for this document.

## B.5 Bursting strength – dry and wet

This test is designed to assess the device's ability to withstand pressure over, for example, a clinician's elbow and to ensure its barrier properties are not prejudiced by mechanical failure.

Materials with more than one layer can show several break points when tested for bursting strength, e.g. one corresponding to each layer. In order to address the scope of the requirement it was agreed to evaluate the performance of the material based on the pressure needed to break or compromise the barrier of the sample.

The limits (Tables 1 and 2) are based on manufacturer's experience of products deemed to be clinically suitable in the market place.

## B.6 Tensile strength – dry and wet

The 'tensile strength' of a material is the maximum stress, generated by pulling or stretching the material that a material can withstand before failing.

The test is designed to assess whether the basic strength of the device material is sufficient to ensure its barrier properties are not prejudiced. It is a standard textile material test.

Materials with more than one layer can show several break points when tested for tensile strength, e.g. one corresponding to each layer. In order to address the scope of the requirement it was agreed to evaluate the performance of the material based on the force needed to break or compromise the barrier of the sample.

The limits (Tables 1 and 2) are based on manufacturer's experience of products deemed to be clinically suitable in the market place.

Tables 1 and 2 have limits for the material in both the wet and dry states, as gowns and drapes are expected to be subjected to wet and dry conditions during use.

## B.7 Resistance to microbial penetration – dry

Dry bacterial penetration EN ISO 22612 is a test method that was designed to simulate the penetration of bacteria-carrying skin scales through fabrics.

This test provides a means for assessing the resistance to penetration through barrier materials of bacteria-carrying particles.

Whilst the relationship between contamination and infection is complex - contamination of the surgical field does not necessarily lead to infection - it is generally agreed that healthcare facilities should consider methods to reduce levels of airborne particles carrying bacteria in operating rooms [27].

The skin is the most important source of airborne contamination in the operating room. A person releases approximately  $10^4$  skin particles per minute when walking and approximately 10 % of these carry bacteria. Activity and friction against the skin, e.g. from clothing, increase the dispersal. When skin scales pass through relatively impermeable clothing, they can also become fragmented, with the result that more than 50 % of the bacteria-carrying particles can be less than 5  $\mu\text{m}$ . Bacteria-carrying skin scales are dispersed from the human body surface mainly from the lower part of the torso.

Normal shedding of human skin cells (keratinocytes) produces individual cells which are approximately 25  $\mu\text{m}$  to 30  $\mu\text{m}$  in diameter (when hydrated) [28]. Whyte and Bailey [29] noted that bacterial-carrying skin scales are on average about 20  $\mu\text{m}$  in size, whilst Mackintosh and colleagues [30] showed that dispersed skin fragments had a wide size range extending below 5  $\mu\text{m}$  for the minimum projected diameter (MPD), with a median MPD about 20  $\mu\text{m}$ , and with 7 % to 10 % less than 10  $\mu\text{m}$ .

The skin scales behave aerodynamically as particles of unit density and size approximately 10  $\mu\text{m}$ . These particles are distributed in the operating room with air currents and settle on exposed surfaces, thereby contaminating the sterile field and causing infection of the surgical site.

For microorganisms to penetrate the material in the dry state, they shall be carried on a physical particle, for example, skin scales. In this test, the physical particles are composed of talcum, where 95 % of the particles shall be  $\leq 15 \mu\text{m}$ . The referenced talcum (Finntalc M15) has a median particle size of 4,5  $\mu\text{m}$ , a maximum size of approximately 17  $\mu\text{m}$ , and approximately 18 % of the particles are  $\leq 2 \mu\text{m}$ .

During the dry penetration test, the talcum particles are sifted through the material to be tested, and spore-forming bacteria are used as marker organisms. The test is intended to measure penetration of dust, e.g. skin scales through clothes, and has been shown to correlate well to airborne dispersal of bacteria.

The size range in the test talcum covers the range of skin fragments found in practice down to particle sizes smaller than we would expect from skin fragmentation.

Penetration in this test method is influenced more by the physical properties of the materials e.g. pore size and tortuosity factor than by their hydrophobic/hydrophilic characteristics.

The limit of  $\leq 300$  CFU (Tables 1 and 2) appears to be partially based on the results of the BIOBAR project<sup>2)</sup> which showed that a standard cotton fabric would allow 1 000 CFU to 10 000 CFU through during the test period that various woven and non-woven laminates allowed no penetration, and that non-woven single-use materials allowed between 150 CFU and 1 000 CFU through. The test is designed to discriminate between materials based on their anticipated particulate penetration properties. Recent tests show that newer materials, both reusable and single-use, are available on the market with lower or no measurable dry penetration.

The decision to only require dry penetration performance for 'less critical product areas' in Tables 1 and 2 is based on agreement in CEN/TC 205/WG 14 that if the critical product area meets the requirements for wet microbial penetration and hydrostatic head, then it will probably also provide resistance against dry microbial penetration. However, the two penetration mechanisms are different and the argument has never been demonstrated.

Dry penetration is intended to examine the ability of a material to prevent airborne transmission. The test is particularly relevant for the clean air suit, which is intended to prevent airborne transmission when made from a tight material and adequately designed.

There are, however, a variety of views on the relevance of airborne transmission for gowns. Whilst there is some evidence that airborne transmission is not prevented by a gown when used in operating

---

2) Project BIOBAR (Contract SMT4-CT96-2123) was funded through the Standards, Measurement and Testing programme, part of the Fourth Framework Programme funded by the European Commission, which investigated test methods for the evaluation of the barrier properties of textile materials against biological infective agents.

rooms with turbulent ventilation [31], [32], many European countries do not use CAS, and therefore there is a body of opinion which believes that good dry penetration properties of gowns are necessary. There is also evidence for the role of gowns in controlling airborne bacterial counts when worn over standard surgical scrubs in operating rooms with laminar vertical downflow ultraclean air ventilation systems [33].

The current requirements in this document are a compromise between these two views.

## B.8 Resistance to microbial penetration – wet

This test determines the resistance of a material to the penetration of bacteria from a dry surface through a material by the combined effect of friction, pressure and wetting [34]. The pressure is intended to mimic the type of pressure exerted by a surgeon's elbow during a procedure [35] and was developed specifically to measure the penetration by bacteria through operation materials of reusable or single-use material.

The method has been difficult to standardize, and multiple inter-laboratory comparisons have taken place where it has been difficult to demonstrate consistent results between laboratories. The method went into immediate revision after initial publication.

As the effects of the modifications of the test protocol on the test results have not yet been investigated and CEN/TC 205/WG 14 has not yet taken a decision on the presentation of results based on the modified test protocol the committee decided to prolong the existing requirements based on EN ISO 22610. CEN/TC 205/WG 14 intends to adapt the performance requirements to the new test protocol as soon as sufficient data are available.

As in former versions of EN 13795-1, the barrier index  $I_B$  is specified to evaluate the conformity of materials with the wet microbial penetration requirements. For critical areas of high performance products, a  $I_B$  of 6,0 is required. 6,0 is the maximum achievable value and means 'no penetration' for the purpose of this document. The requirements for critical areas of standard performance products have been agreed at lower level to anticipate the lower performance level.

The decision not to require wet microbial penetration performance for less critical product areas is based on experts' opinion that a hydrostatic head of 10 cm offered sufficient resistance in these areas and reduced the requirement for extra wet microbial penetration testing. In addition, the pressure on less critical areas is lower, and the risks of strike-through of blood [36] and microbes are also reduced.

## B.9 Labelling

The Medical Device Directive allows manufacturers to use and explain symbols in their instructions for use. In principle experts regarded specifying a uniform set of instructions or symbols which would cover, for example, how to use drapes, as being a benefit for users when using different products. However, such a specification has not yet been developed und hence not included in this document. As labelling requirements are adequately covered in Section 13 of Annex I (Essential Requirements) of the Medical Device Directive the experts found no or only very little need to further specify the Essential Requirements in this document.

## B.10 Treatment of results

The Median,  $M_d$ , was chosen as the preferred statistic to the Mean because of the small sample size and its greater robustness to the influence of outliers. As a consequence, 25th and 75th percentiles ( $L_q$  and  $U_q$  respectively) were chosen as the test statistics for assessing compliance against the performance requirements in Tables 1 and 2. More simply, for  $PR_{\min}$ , for five replicates the highest four shall pass and for 10 replicates the highest eight shall pass. The method for determining  $L_q$  and  $U_q$  in A.3 gives the statistical justification for this.

It was recognized that manufacturers and processors may wish to use means and standard deviations for quality assurance purposes, especially where more data would be generated leading to better estimates of population statistics and the more reliable setting of processing conditions.

## Annex C (informative)

### Information on further characteristics

#### C.1 Comfort

The concept of comfort is based on several different factors, such as physiological comfort, ease of movement or factors that will influence and/or affect the individual's satisfaction with the product.

The thermophysiological comfort of a product depends on such properties as its thermal resistance, air permeability, water-vapour resistance, drapeability, tactile comfort and other properties like stretchability, weight, size, fit, fibres and manufacture.

NOTE 1 Drapeability addresses the ability of a material to conform to a given shape or object.

NOTE 2 Water-vapour resistance is defined as the water-vapour pressure difference between the two faces of a material divided by the resultant evaporative heat flux per unit area in the direction of the gradient. The evaporative heat flux can consist of both diffusive and convective components. EN ISO 11092 provides a test method for measuring the thermal and water-vapour resistance under steady-state conditions.

NOTE 3 Thermal resistance is a property of a material that can be measured by a thermal manikin in view to determine important parameters relevant to clothing thermal comfort.

NOTE 4 Tactile comfort also indicated as softness, is highly dependent on the fibre smoothness and the finish technologies.

NOTE 5 Properties such as stretchability, size fit, weight, can be measured.

Discomfort properties, such as rustling tendency, softness and skin irritation are difficult to measure. Evaluation should be based on trials of the products or practical experience.

#### C.2 Adhesion for fixation for the purpose of wound isolation

Adhesives are used to attach materials during the preparation for an operation and to attach drapes to a patient on the operating table. Different adhesives are chosen for different materials, e.g. material to material and material to the skin.

In choosing an adhesive, the following considerations should be taken into account:

- a) Adhesives should not cause damage to the skin.
- b) When used on reusable materials, the adhesives should be removable during processing without damaging the material.
- c) The adhesive should create a seal-off from liquid and secure a sterile field.

#### C.3 Liquid control

The control of liquids, like body liquids or other liquids used or generated close to the wound during a surgical procedure, is regarded relevant to reduce the risk of transfer of infective agents.

Liquid control can be achieved by several means. Examples of test methods are given in the bibliography but it is regarded as technically impossible to specify a single test method, which addresses all aspects of liquid control and provides comparable results.

#### **C.4 Flammability**

Though surgical gowns and drapes do not provide ignition sources or oxidizer both products might serve as fuel, when a fire breaks out. Manufactures are required to supply information regarding fire risks in relation to the use of their products. This document does not specify further essential requirements of Directive 93/42/EEC on Medical Devices or basic health and safety requirements of the Directive 89/686/EEC on Personal Protective Equipment regarding flammability of surgical gowns and drapes.

#### **C.5 Electrostatic discharge**

CEN/TC 205/WG 14 discussed whether specific tests for Electrostatic Discharge (ESD) were necessary in this document.

After taking advice from clinicians, hospital engineers, experts in electromedical equipment and electrostatic engineers, WG14 note the following:

a) There are three potential risks from ESD:

ESD damage to equipment;

ESD ignition of flammable anaesthetic agents;

ESD ignition of flammable vapours (specifically alcohols).

b) The electrostatic immunity requirement in IEC 60601-1-2:2014 is 15 kV. EN 61000-4-2:2009 has a useful graph in informative Annex A showing that synthetic fabrics can generate a maximum electrostatic voltage of 13 kV in rooms without humidity control (down to 15 %RH). Therefore medical electrical equipment comply to the latest version of EN 60601-1-2 should be adequately protected from ESD.

c) Traditional risks associated with flammable anaesthetic agents no longer exist in hospitals as these agents have all been replaced with safer alternatives.

d) Use of flammable liquids in theatres is controlled, as diathermy would not be viable if there were a risk from sparks. Diathermy is a much greater risk than ESD.

Nowadays, the theoretical risks from ESD therefore appear low.

In addition, CEN/TC 205/WG 14 is unaware of actual reports of patient safety related incidents from ESD, and in the absence of such evidence believes there is no requirement to include ESD testing for gowns and drapes in this document.

CEN/TC 205/WG 14 notes that there are user comfort issues associated with static charge and ESD, and manufacturers can wish to take this into account when selecting materials and designing devices.

**Annex D**  
(informative)

**Environmental aspects**

This document provides test methods and performance requirements for the characteristics of surgical gowns and drapes which enable assessment of compliance with the relevant essential requirements of Annex I of Directive 93/42/EEC on Medical Devices.

In order to reflect the broad variety of technologies currently used to manufacture and (if applicable) process surgical textiles and not to hinder technical development and innovation, the requirements set by this document are expressed in terms of quantifiable performance rather than specific technical design or descriptive characteristics.

This document does not specify any technical solution to meet the requirements set in this document and hence does not include any technical provisions for manufacturing and processing and their respective environmental impacts. As this document does not include environmental provisions, this document does not provide an environmental checklist.



## Annex E (informative)

### Guidance to users for selecting products

#### E.1 Performance levels

This document introduces two performance levels ('standard performance' and 'high performance') for surgical gowns and drapes, thereby acknowledging the fact that products are challenged to differing extents during surgical procedures, dependent upon the duration, mechanical stress and liquid challenge throughout the surgical procedure. The differentiation of 'standard performance' from 'high performance' products is based on the barrier performance of the products in critical product areas.

NOTE 1 For details of the differences in the required barrier performance, see Tables 1 and 2.

By establishing two performance classes this document facilitates the assessment of the barrier performance of products. However, this document does not include specific recommendations for selecting surgical gowns or drapes with regard to the type of surgical procedure the product is to be used with.

The user will select surgical gowns and drapes based on their performance in order to meet the anticipated challenges of the surgical procedure (e.g. in terms of duration, mechanical stress and liquids). If the classification scheme provided by this document is not considered suitable to address the anticipated challenges during use, discrete test results for the characteristics to be evaluated can be taken as a basis for selecting products.

NOTE 2 The selection and use of surgical gowns and drapes for specific surgical procedures can be covered by risk assessment and quality management carried out by the user and can be subject to local, regional or national infection prevention regime, guidelines, directives or regulation.

#### E.2 Functional design

##### E.2.1 General

This document does not include specific requirements for the functional design of surgical gowns and drapes. The impact of functional design on the performance of products is acknowledged by requiring testing on the finished product including potential weak spots.

However, the functional design – in particular critical and less critical areas, the over-all size of the product and the characteristics of accessories (if any) – and its impact on the working situation (thermophysiological comfort and ergonomics) should be considered when selecting products for use.

##### E.2.2 Critical and less critical areas

This document acknowledges the fact that not all areas of the product are involved in the transfer of infective agents to or from the wound to the same extent. In order to set different performance requirements and allow for different product areas this document introduces 'critical product areas' and 'less critical product areas'.

NOTE 1 In general 'critical product areas' include those areas most likely to be exposed to blood and other body liquids as, e.g. front and sleeves of surgical gowns or the parts of surgical drapes adjacent to the surgical wound. The back of a surgical gown and part of surgical drapes being far from the wound are usually considered as 'less critical product areas'.

NOTE 2 For details of the differences in the required performance of 'critical product areas' and 'less critical product areas', see Tables 1 and 2.

This document does not include provisions for the size and position of 'critical' or 'less critical' product areas. The user has to decide whether or not size and position of 'critical' and 'less critical' product areas are suitable to meet the anticipated challenges of a certain surgical procedure.

### E.2.3 Size

This document does not include provisions for specifying the size of products in a standardized way.

Selecting products of suitable size in order to appropriately cover persons, patients and equipment is up to the user in order to ensure the intended use of the respective product.

NOTE Using products of inappropriate size might lead to insufficient covering, i.e. jeopardize the aim of minimizing the transfer of infective agents, and might impact freedom or safety of movements (e.g. with gowns too small or too big for the wearer).

### E.2.4 Accessories

This document does not include specific provisions for accessories such as, e.g. cuffs or buttons.

As accessories do therefore not need to meet any requirements of this document, the user should assess the functional design with consideration to the placement of accessories so that the intended uses of the products are not compromised. The user should also assess the quality of any accessories in order to ensure that the intended uses of the products are not compromised.

### E.2.5 Comfort

#### E.2.5.1 General

The functional design of products has an impact on the thermophysiological comfort.

NOTE 1 For more information on comfort, see C.1.

The user when selecting products for use should assess the comfort of products in order to exclude any significant limitations of the intended use of the product. Combinations of materials and design of clothing systems (including technical underwear or garments) that will minimize the physiological stress during work are to be encouraged.

NOTE 2 The comfort of surgical gowns and drapes depends on various characteristics, most of which can be evaluated using standardized test methods. More easily the overall comfort of surgical gowns and drapes can be assessed with trials (i.e. personal experience).

#### E.2.5.2 Surgical gowns

The overall comfort of surgical gowns can be influenced by a number of factors: design, fit, breathability, weight, surface thickness, electrostatic properties, colour, light reflectance, odour and skin sensitivity.

Other important variables that can influence comfort include undergarments, health and physical conditions, workload, mental stress and environmental conditions, such as temperature, relative humidity, and air changes in operating room.

The perception of comfort is subjective and can be influenced by one or a combination of the aforementioned factors.

### **E.2.5.3 Surgical drapes**

Surgical drapes should be flexible so that they will cover the patient closely and smoothly, allowing placement and manipulation of instruments and draping of other related equipment, such as ring stands, back tables, and Mayo stands.

Liquid control is important for surgical drapes in operations with much blood or other liquids such as saline.

### **E.3 Practical trials**

Not all the necessary properties of a product can be tested according to this document. The products should be tested practically in clinical situations where the end-user is going to apply them, to ensure that they are suitable from all important aspects including functionality and comfort. The practical trials should be evaluated before choice of products.

**Annex ZA**  
(informative)

**Relationship between this European standard and the essential requirements of Directive 93/42/EEC [1993 OJ L 169] aimed to be covered**

This European Standard has been prepared under a Commission's standardization request 'M/295 concerning the development of European standards related to medical devices' to provide one voluntary means of conforming to essential requirements of Council Directive 93/42/EEC of 14 June 1993 concerning medical devices [1993 OJ L 169].

Once this standard is cited in the Official Journal of the European Union under that Directive, compliance with the normative clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding essential requirements of that Directive, and associated EFTA regulations.

NOTE 1 Where a reference from a clause of this standard to the risk management process is made, the risk management process needs to be in compliance with Directive 93/42/EEC as amended by 2007/47/EC. This means that risks have to be reduced 'as far as possible', 'to a minimum', 'to the lowest possible level', 'minimized' or 'removed', according to the wording of the corresponding essential requirement.

NOTE 2 The manufacturer's policy for determining **acceptable risk** must be in compliance with Essential Requirements 1, 2, 5, 6, 7, 8, 9, 11 and 12 of the Directive.

NOTE 3 When an Essential Requirement does not appear in Table ZA.1, it means that it is not addressed by this European Standard.

**Table ZA.1 — Correspondence between this European standard and Annex I of Directive 93/42/EEC [1993 OJ L 169]**

Essential Requirements of Directive 93/42/EEC	Clause(s) / subclause(s) of this EN	Remarks / Notes
7.3, first part only	5.2, A.2.7	Covered for wet microbial penetration.
8.1, first sentence only	4, A.2.1, A.2.6, A.2.7	Covered for cleanliness (bioburden), and dry and wet microbial penetration.
9.2, second indent only	4, A.2.4 and A.2.5	Covered for bursting strength (dry and wet) and tensile strength (dry and wet).
13.6 h), first paragraph only	6.2.1, 6.2.2	Covered for the number of reuses and the sterilisation process to be applied for products to be terminally sterilised.
13.6 i)	6.2.1, 6.2.2	Covered for the sterilisation process to be applied for products to be terminally sterilised.

**WARNING 1** — Presumption of conformity stays valid only as long as a reference to this European standard is maintained in the list published in the Official Journal of the European Union. Users of this standard should consult frequently the latest list published in the Official Journal of the European Union.

**WARNING 2** — Other Union legislation may be applicable to the product(s) falling within the scope of this standard.

## Bibliography

- [1] EN 1041, *Information supplied by the manufacturer of medical devices*
- [2] EN 14065, *Textiles - Laundry processed textiles - Biocontamination control system*
- [3] EN 62366, *Medical devices - Application of usability engineering to medical devices*
- [4] EN ISO 9073-6, *Textiles - Test methods for nonwovens - Part 6: Absorption (ISO 9073-6)*
- [5] EN ISO 9073-11, *Textiles - Test methods for nonwovens - Part 11: Run-off (ISO 9073-11)*
- [6] EN ISO 9073-12, *Textiles - Test methods for nonwovens - Part 12: Demand absorbency (ISO 9073-12)*
- [7] EN ISO 9237, *Textiles - Determination of permeability of fabrics to air (ISO 9237)*
- [8] EN ISO 10993-5, *Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity (ISO 10993-5)*
- [9] EN ISO 10993-10, *Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization (ISO 10993-10)*
- [10] EN ISO 11092, *Textiles - Physiological effects - Measurement of thermal and water-vapour resistance under steady-state conditions (sweating guarded-hotplate test) (ISO 11092)*
- [11] EN ISO 11607-1, *Packaging for terminally sterilized medical devices - Part 1: Requirements for materials, sterile barrier systems and packaging systems (ISO 11607-1)*
- [12] EN ISO 11607-2, *Packaging for terminally sterilized medical devices - Part 2: Validation requirements for forming, sealing and assembly processes (ISO 11607-2)*
- [13] EN ISO 11810, *Lasers and laser-related equipment - Test method and classification for the laser resistance of surgical drapes and/or patient protective covers - Primary ignition, penetration, flame spread and secondary ignition (ISO 11810)*
- [14] EN ISO 13485, *Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485)*
- [15] EN ISO 15223-1, *Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements (ISO 15223-1)*
- [16] EN ISO 15797, *Textiles - Industrial washing and finishing procedures for testing of workwear (ISO 15797)*
- [17] Council Directive 93/42/EEC of 14 June 1993 concerning medical devices, OJ L 169, 12.7.1993, p. 1-43
- [18] VON WOEDTKE T. The limits of sterility assurance. *GMS Krankenhhyg. Interdiszip.* 2008 Sep 3, 3 (3) p. Doc19. Available at: Kramer A
- [19] Laird J, et al. Foreign body contamination during Interventional procedures, An underrecognized factor that can cause catastrophic complications. *Endovascular Today* Sept 2012;94-97

- [20] TRUSCOTT W. Impact of microscopic foreign debris on post-surgical complications. *Surg. Technol. Int.* 2004, **12** pp. 34–46 [Review]
- [21] JANOFF K., WAYNE R., HUNTWORK B., KELLEY H., ALBERTY R. Foreign Body reactions Secondary to Cellulose Lint Fibres. *Am. J. Surg.* 1984, **147** pp. 598–600
- [22] TINKER M.A., TEICHER I., BURDMAN D. Cellulose granulomas and their relationship to intestinal obstruction. *Am. J. Surg.* 1977 Jan, **133** (1) pp. 134–139
- [23] CINA S.J., RASO D.S., CRYMES L.W., UPSHUR J.K. Fatal suture embolism to the left anterior descending coronary artery. A case report and review of the literature. *Am. J. Forensic Med. Pathol.* 1994 Jun, **15** (2) pp. 142–145 [Review]
- [24] MAKINO K., YAMAMOTO N., HIGUCHI K., HARADA N., OHSHIMA H., TERADA H. Phagocytic uptake of polystyrene microspheres by alveolar macrophages: effects of the size and surface properties of the microspheres. *Colloids Surf. B Biointerfaces.* 2003, **27** p. 3339
- [25] ELEK S.D., CONEN P.E. The virulence of *Staphylococcus pyogenes* for man; a study of the problems of wound infection. *Br. J. Exp. Pathol.* 1957 Dec, **38** (6) pp. 573–586
- [26] JR EDMISTON C.E., SINISKI S., SEABROOK G.R., SIMONS D., GOHEEN M.P. Airborne particulates in the OR environment. *AORN J.* 1999, **69** (6) pp. 1169–1172, 1175–1177, 1179
- [27] A. Hambraeus. Aerobiology in the operating room – a review *J hosp Inf* 1988; 11 (suppl A): 68-76
- [28] NOBLE W.C., LIDWELL O.M., KINGSTON D. The size distribution of airborne particles carrying microorganisms. *J. Hyg. (Lond.)*. 1963, **61** (4) pp. 385–391
- [29] WHYTE W., BAILEY P.V. Reduction of microbial dispersion by clothing. *J. Parenter. Sci. Technol.* 1985 Jan-Feb, **39** (1) pp. 51–61
- [30] MACKINTOSH C.A., LIDWELL O.M., TOWERS A.G., MARPLES R.R. The dimensions of skin fragments dispersed into the air during activity. *J. Hyg. (Lond.)*. 1978, **81** (3) pp. 471–479
- [31] HAMBRAEUS A., LAURELL G. Protection of the patient in the operating suite. *J. Hosp. Infect.* 1980, **1** pp. 15–30
- [32] PASQUARELLA C., PIZURRA O., HERREN T., POLETTI L., SAVINO A. Lack of influence of body exhaust gowns on aerobic bacterial surface counts in a mixed-ventilation operating theatre. A study of 62 hip arthroplasties. *J. Hosp. Infect.* 2003, **54** (1) pp. 2–9
- [33] WHYTE W., VESLEY D., HODGSON R. Bacterial dispersion in relation to operating room clothing. *J. Hyg. (Lond.)*. 1976 Jun, **76** (3) pp. 367–378
- [34] SMITH J.W., TATE W.A., YAZDANI S., GARACIA R.Y., MUZIK A.C., NICHOLS R.L. Determination of surgeon-generated gown pressures during various surgical procedures in the operating room. *Am. J. Infect. Control.* 1995 Aug, **23** (4) pp. 237–246
- [35] RANSJÖ U., HAMBRAEUS A. An instrument for measuring the bacterial penetration through fabrics used for barrier clothing. *J Hyg Camb.* 1979, **82** pp. 361–368
- [36] BELKIN N.L. Strike-through, gown safety. *AORN J.* 1992 Oct, **56** (4) pp. 627–628